Specification Sheet

Product: EALI515-T (Code# 1502, 1512, 1513, 1521)

EXAMINATION	METHOD	SPECIFICATION
Color	Visual test	Orange - Red
Turbidity	Visual test	Clear
pH (After equilibration to 5% CO_2 at 37°C)	JP* standard	7.4-7.7
Osmotic pressure	Freezing point depression method	290±10 mOsm/kg H_2O
Endotoxin	Kinetic-chromogenic technique by JP* standard	< 0.1EU/ml
Sterility	Sterility test by JP* standard	Negative
Mycoplasma	Nucleic acid amplification test	Negative
Cell culture	Cell proliferation test by HPBL**	Company Standard

* Japanese Pharmacopoeia
** Human peripheral blood lymphocyte